

Code 2981

50 Test Strips

Low Range
Hardness



InstaTest[®] ANALYTIC

Low Range Hardness ppm (ppg)



Using a cup-size sample, immerse strip and remove with **pad face up**. **DO NOT SHAKE OFF EXCESS WATER**. Wait 15 seconds and compare to color chart. Discard strip after 30 seconds.



- Keep wet fingers out of the bottle.
- Cap bottle tightly.
- Store in a cool, dry place.



LaMotte Company
Chestertown, Maryland 21620 USA
<http://www.lamotte.com>



6 373951466819