



## BROMINE KIT


OCTA-SLIDE 2, 0.2-3.0 ppm  
CODE 6955-01

QUANTITY	CONTENTS	CODE
50	*Bromine Tablets	*6977A-H
2	Test Tubes, plastic, 2.5-5-10mL, w/caps	0106
1	Octa-Slide 2 Viewer	1101
1	Bromine Octa-Slide 2 Bar, 0.2-3.0 ppm	6822-01

\*WARNING: Reagents marked with an \* are considered to be potential health hazards. To view or print a Material Safety Data Sheet (MSDS) for these reagents go to [www.lamotte.com](http://www.lamotte.com). To obtain a printed copy, contact LaMotte by e-mail, phone or fax.

To order individual reagents or test kit components, use the specified code number.







### USE OF THE OCTA-SLIDE 2 VIEWER



The Octa-Slide 2 Viewer should be held so non-direct light enters through the back of the Viewer. Slide the Octa-Slide 2 Bar into the Viewer. Insert the reacted sample into the top of the Viewer. Match the color of the reaction to the color standards.

**WARNING!** This set contains chemicals that may be harmful if misused. Read cautions on individual containers carefully. Not to be used by children except under adult supervision.

## PROCEDURES

<p><b>1</b></p>  <p>Insert Bromine Octa-Slide 2 Bar (6822-01) into the Octa-Slide 2 Viewer (1101).</p>	<p><b>2</b></p>  <p>Fill a test tube (0106) to the 5 mL line with the water sample.</p>	<p><b>3</b></p>  <p>Add one *Bromine Tablet (6977A).</p>
<p><b>4</b></p>  <p>Cap and mix until tablet disintegrates.</p>	<p><b>5</b></p>  <p>Insert test tube into Octa-Slide 2 Viewer.</p>	<p><b>6</b></p>  <p>Match sample color to a color standard. Record as ppm Bromine.</p>

**NOTE:** Thoroughly clean and rinse test tubes after each use.

### LaMOTTE COMPANY

Helping People Solve Analytical Challenges®

PO Box 329 • Chestertown • Maryland • 21620 • USA  
800-344-3100 • 410-778-3100 (Outside U.S.A.) • Fax 410-778-6394

Visit us on the web at [www.lamotte.com](http://www.lamotte.com)