

Pack content:

1 aluminium container with 100 test strips

Measuring range:

Visually
10–1000 mg/L SO_3^{2-}

Reflectometrically
10–500 mg/L SO_3^{2-}

Color gradation:

0 · 10 · 25 · 50 · 100 · 250 · 500 · 1000 mg/L SO_3^{2-}

General indications:

Remove only as many test strips as are required. Close the container immediately after removing a strip. Do not touch the test field.

Instructions for use:

1. Dip the test strip into the sample (pH 6–9) for 1 s.
 2. Shake off excess liquid.
 3. Wait 20 s.
 4. Compare the test field with the color scale.
- If sulfite ions are present, the test field turns salmon-pink.

Interferences:

Acidic test solutions (i.e. white wine) must be buffered with solid sodium acetate prior to testing because the test paper does not react in acidic media. Sulfide ions also lead to a red coloration of the test paper in neutral solutions. This interference can be eliminated by adding an excess of nickel (II) ions. The nickel sulphide formed can be removed by filtration. Larger amounts of reducing agents such as hydroxylammonium chloride or ascorbic acid interfere with the determination leading to falsely low test results.


Disposal:

Used test strips can be placed in the normal household waste.

Storage:

Avoid exposing the strips to sunlight and moisture. Keep container cool and dry (storage temperature not above +30 °C).

If correctly stored, the test strips may be used until the use-by-date printed on the packaging.



REF	Artikelnummer / Item number / Référence produit / Referencia / Artikelnummer / Numero di catalogo / Numer artykułu	LOT	Chargenode / Batch identification / Número de lot / Código de lote / Productienummer / Codice del lotto / Numer partii
	Achtung! / Warning! / Attention! / ¡Atención! / Let op! / Attenzione! / Uwaga!		Verwendbar bis / Use by / À utiliser avant / Fecha de caducidad / Te gebruiken tot / Utilizzare entro (anno / mese) / Przydatność do użycia
	Temperaturbegrenzung / Permitted storage temperature range / Limites de température / Límites de temperatura / Temperaturbegrenzung / Limite de temperatura / Ograniczenie temperatury		Trocken aufbewahren / Store in a dry place / Conserver au sec / Mantener seco / Op een droge plaats bewaren / Mantener asciutto / Przechowywać w suchym miejscu
	Gebrauchsanweisung beachten / Please read instructions for use / Respecter les instructions d'utilisation / Obsérvense las instrucciones de uso / Lees de bijsluiter / Consultare le istru- zioni per l'uso / Przestrzegać instrukcję użycia		Packung geschlossen halten / Keep container closed / Refermer la boîte / Mantenga el envase cerrado / Verpak- king gesloten houden / Conservare la confezione chiusa / Przechowywać pojemnik szczelnie zamknięty
	Ausreichend für <n> Prüfungen / Contains sufficient for <n> tests / Contenu suffisant pour <n> tests / Contenido suficiente para <n> tests / Voldoende voor <n> tests / Contenuto sufficiente per <n> test / Wystarczająco dla kontroli <n>		